



DIRECTORATE OF ANIMAL HUSBANDRY & VETERINARY SERVICES, ODISHA, MANGALABAG, CUTTACK-753001  
Phone No.0671-2414629, Fax No.0671-2414739, E-mail: dahvsorissa@gmail.com

No. 2695 /Vet. Dt. 27.02.2024

**PRE-BID CORRIGENDUM**

**FOR SUPPLY OF VETERINARY CARE EQUIPMENT, CONSUMABLES AND ANCILLARIES**

**Bid Ref. No-01/2023-24/Veterinary care equipment, consumables and ancillaries dt.01.02.2024**

Sl. No	Item Name/Particulars	Relevant specification as mentioned in the Tender Document against which queries has been raised	Queries raised by the prospective Bidder	Clarification/ Amendments in response to the queries
	<b>Technical Specification</b>			
1	<b>Item Sl.No.32 :Automated Nucleic Acid Extraction System</b>	Liquid handling Parameters: The instrument should have facility to adapt the aspiration dispensing parameters as per liquid viscosity	Technical specification to be modified as per Magnetic based Technology	No Change
2	<b>Item Sl.No.39 :Flow Cytometer</b>	1. Software Functionality and assays: Instrument QC Guava Check TM software for evaluation and tracking of instrument performance 2.Multi Colour Detection Assays Guava Express Plus TM Express Pro,insteadRed,green, Yellow detection Channels/Semen Analysis Software CytoSoft,IMV basedsoftware for sample analysis with pre-defined settings per species & assays.	Flowcytometer appears to be specific to a particular Company doesn't give fair chance for participating in the tender.	No Change
3	<b>Item Sl.No.40 :Mini AI Gun Caprine for Cervical/Trans Cervical artificialInsemination</b>	AI gun for Goat	Gun type universalO Ring or conical O Ring	Amended: Gun type: Universal O Ring, Other specification as per bid document shall remain unchanged
4	<b>Item Sl.No.35: PPE Kit</b>	House gloves- Disposable ISI Mark Shoe Cover: Disposable with thick padded foot base	House gloves with ISI Mark surgical or examination Shoe Cover:Type of disposable material whether Plastic/woven	Amended: House gloves with ISI Mark for examination Shoe Cover material: Plastic or wooven Other specification as per bid document shall remain unchanged
5	<b>Item Sl.No.48: Video Assisted Reproductive Tract Diagnostic tool cum Digital Artificial Insemination Gun</b>	Mobile Specification OS Android 12.0, ram.4-6GB, Connectivity technology Bluetooth, WIFI, USB	OS of Mobile phone to be supplied in the transport case Whether iOS or Android	Operating System of the Mobile phone to be supplied in the transport case should have Android 12.0.

				Hence NO CHANGE
6	Item Sl.No.78 : Bench Platform Stainless Steel Weighing Balance(Capacity 200 Kg)	Calibration-Internal,Weighing Capacity up to 200 Kg,Material-MS	1. Type of Material Stainless Steel or MS 2.Size of the Platform:	Amended:1. Material Stainless Steel, 2. Size of Platform: 60cm X 60 cm Other specification as per bid document shall remain unchanged
7	Item Sl.No.79 : Absorbent Cotton IP 500 gmPkt(Each)	500 Gm, Absorbency not more than 10sec	Gross weight or Net Weight	Amended: Net Weight 500 gm. Other specification as per bid document shall remain unchanged
8	Item Sl.No.80:Cryovial 1.8 ml/4.5 ml.PPpkt of 100	Cryovial 1.8 ml/4.5 ml.PP PKT OF 100. Material -plastic	Cryovial 1.8 ml or Cryovial 4.5 ml. Any other information about the lock system of cryovial	Amended: Cryovial 1.8 ml with internal O ring and thread system. Other specification as per bid document shall remain unchanged
9	Item serial No-27-Pipette (two sets) each 1-2.5 ul, 1-10 ul, 10-100 ul, 10-200 ul, 1000 ul	Spring Loaded Tip Cone: Improved ergonomics, Tight fit to the tip (not in 5ml and 10ml pipettes)	Spring loaded tip cone manufactured by one German company only. Universal type is manufactured by many manufacturers. Requesting to change the specification	No change
10	ITEM No-10 -Digital LCD Microscope, ITEM No-24- U.V. Spectrophotometer , Item No-29-Elisha reader, Item No-32 Automated nucleic Acid extraction system, Item No-39 ,Flow cytometer	Specification as per bid document	The specification of the items appears to be specific to a particular vender	No change
11	ItemNo-- 1,2,8,9,10,13,14,15,16,17,18,19 ,20,21,22,23,24,25,26,27,28,29, 30,31,32,33,38,42,48,49,50,58, 59,60,61,64,67,72,80	Specification as per bid document	Specification should be general to get best competitive price without scarifying quality to avoid fixing of tender by way of lock in specification.	No change
12	Item Sl.No.64 : Portable Veterinary Ultrasound Machine	3. It should have intuitive interface, touch pad and touch key board 4. It should have Touch system to select imaging modes like: B, B+B, Power B, B+M, M, CFM and PW. 8. It should have minimum 2 USB connectors, wi-fi compatibility to send images to iPad and other smart devices , Basic DICOM, stereo sound and Bluetooth compatibility .	3.Modification sought for intuitive interface, touchpad/track ball and touch key board 4.Touch screen or control panel key/ botton to select imaging modes like B,B+B, power B, B+M, M, CFM, PW 8.Minimum 2 USB connectors, Dicom inter compatibility and	No change



		<p>15. The machine should be having HDMI and wi-fi compatibility</p> <p>17. The Machine must be able to take the voice command of the user to adjust depth, probe frequency, focus, gain and freeze in B mode</p> <p>18. It should be provided with one Micro Convex probe having 5-7.5Mhz frequency suitable for small animal abdomen.</p> <p>19. It should be offered with Endo Rectal probe having frequency of 5-10Mhz and 600 mm footprint of the scanning face of the probe for large animal reproduction and fetal sexing on animal. The Endo rectal probe should be at least 2.5 meter long.</p>	<p>Dicom to PACS</p> <p>15.Having HDMI and LAN compatibility</p> <p>17. Remove this point</p> <p>19.Endo rectal probe 13mm or more foot print</p>	
13	<b>Item Sl.No.8: Portable digital Radiography</b>	<p>High frequency microprocessor controlled 2.5 KW 40Khz</p> <p>40 to 100 KVP with an increment of 1 KVP per step. 45 mA.</p> <p><u>FLAT PANEL DETECTOR (Wi fi)</u></p> <p>Stationary anode X-Ray tube with focal spots of 1.2mm 14 kJ 2.8K x 2.3K 3.33 lp/ mm 150µm Less than 4kgs</p>	<p>High frequency microprocessor controlled 1.6KW or more</p> <p>40 to 90 KVP with an increment of 1 KVP per step. 40 mA</p> <p>Stationary anode X-Ray tube with focal spots of 1.8mm or better 3K x 4K or more 5 lp/ mm 100µm Less than 2.5kgs</p>	No change
14	<b>Terms &amp; Conditions</b>			
15	<b>Clause 5.1.1 Delivery period</b>	60 days from date of issuance of Purchase Order.	In respect of Items to be imported, it is difficult to make successful execution of supply of products within 60 days. So requesting to grant minimum 90 days delivery schedule.	Amended-90 days from date of issuance of Purchase Order
16	<b>Clause 5.1.1 Delivery period</b>	60 days from date of issuance of Purchase Order.	Deliver Schedule for X-Ray and Ultra sound : 60 days after AERB and PNDT permission	Amended-: 60 days after AERB and PNDT permission
17	Section -IV, Clause-4 of Important Notes:	The bidders shall have to quote the price of AMC (excluding GST) (for the items mentioned) in the prices schedule in the price BOQ (attached as excel file) in the e-tender portal.	AMC price. Separate column for AMC in BOQ	Amended-As per Clause-6.39.1 The decision to enter into AMC will be determined based on cost and complexity of the equipment by the Tender Inviting Authority or User

				Institution, at its discretion, prior to the expiration of warranty period. Thus bidders are not required to quote the AMC in the price BOQ.
18	<b>Clause 5.2.5.2 Experience</b>	The Bidder must have 3 (three) years' experience in supplying Veterinary equipment and instrument, of its own manufacturing or reputed manufacturers of National level to <b>Central &amp; State Government</b> / semi Govt. organisations, PSUs, in the State of Odisha with annual average worth of Rs 05 (Five) Cr. or more in the any three (3) consecutive financial years during 2018-19, 2019-20, 2020-21, 2021-22 & 2022-23.	<p>Veterinary instruments &amp; equipment should be amended as Lab. instrument &amp; equipment and average annual turnover should be Rs.1.00 crore instead of Rs.5.00 crore</p> <p>Exemption from experience in supply of veterinary items for MSME and Start ups</p>	<p>Amended: Veterinary instruments &amp; equipment shall be read as instrument&amp; equipment used in Veterinary Field/Laboratory. Other specification as per bid document shall remain unchanged</p> <p>No change</p>
19	<b>Clause 6.9.1 EMD</b>	Bidders except MSE & start-ups are required to deposit EMD of Rs.1,50,00,000/- (Rupees Onecrore fifty lakhs) as per Rule 212(1) of OGFR 2023 as bid security.	<p>To fix-up EMD against individual items or by segregating items in groups</p> <p>Instead of Rs 1.5 Cr same should be amended as per previous tender</p> <p>Adjustment of EMD for Performance security</p>	<p>No Change</p> <p>No Change</p> <p>No Change</p>
20	<b>Clause 5.2.5.1 &amp; 5.2.6.2 Annual Average turnover</b>	Proof of annual average turnover (Manufacturers/Importer) of Rs.6 (Six)Crores or more in the any three (3) consecutive financial years during 2018-19, 2019-20, 2020-21, 2021-22& 2022-23certified by the Chartered Accountant	<p>To increase the turn over to Rs.15.00 Crore to get genuine and guaranteed bidder</p> <p>The average turnover should be Rs.1.00 crore</p> <p>Turn over exemption for Manufacturer under MSME and Start ups</p>	<p>No Change</p> <p>No Change</p> <p>No Change</p>
21	<b>Clause 6.22.1</b>	Copy of ADHAAR	There is no provision to issue Aadhar in the name of any company	Aadhar is not mandatory for company
22	<b>Pre-bid Meeting SI No-3 , DTCN</b>	On 08.02.2024 at 3.00PM through online mode	<p>Requesting for offline pre-bid meeting for more clarity</p> <p>Since the suppliers are from outside the country it is not possible to attend the offline meeting. Thus requesting to for virtual pre-bid meeting</p>	No Change



**Clarification as sought for by the bidders:**

**6.29.2** - Price Offered shall be in Indian Rupees and inclusive of Price for the supply, installation, training (if necessary) and successful commissioning of the accessories and fulfillment of warranty/guarantee and after sales service to the satisfaction of the User Institution.

**6.29.6** – Self explanatory.

**6.29.8** – Self explanatory.

**6.29.10** – There is no column for AMC in the BOQ format . Please refer Clause 6.39.1.

**6.29.11** - The basic price shall not include the AMC. It will be decided by the Tender Inviting Authority prior to the expiration of warranty period.

**6.29.12** – There is no column for AMC in the BOQ format. The bidders are to quote the price for instrument.

**6.29.13** – The bidders are to quote the price for instrument not for its consumables .

**6.30 Financial Bid Evaluation**

**6.30.2** As per Clause 6.39.1

**6.36 Payment**

**6.36.4** As per Clause 6.39.1

**6.37.1** As per Clause 6.39.1

**6.37.3** Self explanatory

**6.37.4** Self explanatory

**6.38. Guarantee/Warranty Terms: Self explanatory**

**6.39 Annual Maintenance Contract (AMC)** - The decision to enter into AMC will be determined based on cost and complexity of the equipment by the Tender Inviting Authority or User Institution, at its discretion, prior to the expiration of warranty period.

**6.40 Spare Parts/Reagents** – The bidders are not required to quote the price for spare parts and reagents.

**General clarification-**

1. The bidders shall not quote Zero price in BOQ file in the e- tender portal, it will not be taken into account for evaluation.
2. Bidders who want to quote for qPCR (Real time) must quote for item numbers 16,17,18,19, 20,21,22,23,24,25,26,27 and 28 to ensure the qPCR setup is complete and accurate. It is the bidder's responsibility to install all the equipment, including qPCR and other related items, in the laboratory. The bidder must ensure that all the equipment is installed correctly and is functional. Additionally, the bidder is responsible for providing training on operation procedures of all the equipment they provide.

Hence, bidders have to quote for all the items mentioned above or else quoted price will not be considered.

3. Except the above mentioned amendments all other specification and terms and condition remain unaltered.

**Corrigendum for extension of Date & Time**

1	End Date & time of Online bid submission	As per Tender Notice	Amended
		29.02.2024 at 05.00 PM	16.03.2024 at 05.00 PM
2	Date & time of online Technical bid opening	02.03.2024 at 11.30 AM	19.03.2024 at 11.30 AM

  
**DIRECTOR**  
Animal Husbandry & Veterinary Services,  
Odisha, Cuttack