

**DIRECTORATE OF ANIMAL HUSBANDRY & VETERINARY SERVICES, ODISHA,
MANGALABAG, CUTTACK-753001.**

Phone No.0671-2414629, Fax No.0671-2414739, E-mail:dahvsorissa@gmail.com

No. 4487 /Vet.

Date 13.03.2025

Pre-Bid Corrigendum

**Bid reference No-01/2024-25/DAHVS/Veterinary Care Equipment, Consumables & Ancillaries
dt.30.01.2025**

Sl. No.	Particulars	Relevant specification as mentioned in the Tender Document against which queries has been raised	Queries raised by the prospective Bidder	Clarifications / Amendments in response to the queries
1	EMD – Section-III Clause 3.1	Bidders except MSE / OSME & START-Ups are required to deposit EMD of Rs.1,00,00,000/- (Rupees One crore) as per Rule 212(1) of OGFR 2023 and vide MSME Department Notification No-566/ MSME dt.24.01.2024 as bid security. The OSME / local MSE & Start-Ups of Odisha are exempted to furnish the EMD. To claim exemption the bidders must have to submit the supporting documents.	EMD should be 2-5% of tender value, but the bid value is not mentioned in the bid document. Requested for provision of individual item wise EMD. Raising of EMD, that will attract committed and financially sound supplier to ensure competitive price and in time delivery.	EMD has been determined between 2 to 5% of the estimated tender value. As per Rule 212(ii) of OGFR-2023, the exact amount of bid security (EMD) should be determined and indicated in the bid document. It is not mentioned in the said rule to declare the tender value and percentage of EMD. The EMD value has been fixed at Rs. 1 Crore which is mentioned in the bid document.
2.	Turnover – Turnover & Experience- Clause-5.2.6.1 & 5.2.6.2	Annual average turnover (Manufacturers/Importer) of Rs.5.50 Cr (Five Crore Fifty Lakh) or more in any three (3) consecutive financial years during 2018-19, 2019-20, 2020-21, 2021-22, 2022-23. Bidder must have 3 (three) years' experience in supplying Veterinary equipment and instruments of its own manufacturing or reputed manufacturers of National level to Central & State Government / semi Govt. organizations, PSUs Annual average worth of Rs.4.50Cr. or more in the any three (3) financial years during 2018-19, 2019-20, 2020-21, 2021-22, 2022-23.	Clarify whether bidder participating for only few products must have to meet the requirement of Turnover & Experience. Request to reduce the annual average amount of turn over & experience. To increase annual turnover to Rs.15.00-25.00 Crore & minimum experience value requirement to Rs.08.00-12.00 Crore to ensure financially sound supplier.	To apply for few items, bidders must have to meet the turn over and experience requirement as per bid document. No change No change

			Requesting to reduce the turnover & experience requirement to 3-4 Crore. Proposed to add a clause stipulating that, at least 20% of the overall experience value must be derived from supplying to Odisha Govt. in order to safeguard the quality and reliability of the product.	No change No change
3.	Rate Contract – Clause-2.2.3	This is a Rate contract Bid, the rate of which will be valid for a period of 2 (Two) years from the date of finalization of rate contract. During the validity period Director, AH & VS, Odisha may place repeated order to supply the same item procured under this tender irrespective of the quantity mentioned in it.	Clarify if the quantity is mentioned in the tender are for 2 year terms or for 1 year.	As mentioned in the bid document, it is a Rate contract tender. Quantity mentioned in the bid is tentative and it will be increased or decreased substantially. The rate derived as result of the tender will remain valid for 2 years.
4.	Delivery & Installation – Clause 5.1.1	Delivery destination point is mentioned in Section-IV of the bid document. Delivery should be made within 70 days from date of issuance of Purchase Order.	Give sufficient time for delivery. Equipment wise destination should be mentioned. Quantity should be mentioned in BoQ.	Delivery time & item wise destination has already been mentioned in Section-IV of the bid document. As it is a Rate Contract bid, unit price of the item has to be quoted by the bidder.
5.	Item Sl. No.70, 71, 72, 73	Quality certificate as per bid – Products Certified with GMP - QMS from FDCA, CE, ISO 13485, ISO 9001, ISO 45001	Omit GMP - QMS from FDCA as this standard syringe is not use in Veterinary field.	No Change
6.	Manufacturer's authorization – Clause- 5.2.7.1	They have to submit manufacturer's authorization from the original equipment manufacturer (OEM) as per Format-T7. Manufacture cannot authorize multiple agents for same product.	Requesting to allow multiple authorized distributor of an OEM to ensure stock availability, continuity of supply and aligning to operational time lines.	OEM can authorize maximum two agents for same product in addition to itself, as this will encourage competitive bidding to achieve best possible price & quality.
7.	Section-IV Portable Veterinary Ultrasound Machine	Specification as per bid document	Request to revise the Technical Specification of Portable Digital USG at Sl. No. 1, 2, 3, 4, 7, 8, 9, 10, 13, 15, 17, 19, 21, 22 & 26.	No change
8.	Section-IV Portable digital Radiography	Specification as per bid document	Revise the technical specifications of Generator & tube, X-Ray tube head, net	No change

			size, net weight, Weight of stand, Quality certificate, portable Digital Radiography and Delete of conditions like imparting training to the operator post installation and Delete of the conditions like "Image aquisition software, X-Ray generator and Flat panel detector should be from same manufacturer for seamless connectivity and consistent image.	
9.	Section-IV Digital/Electronic semen thaw unit 475 ml capacity	Specification as per bid document	Additional specification like "Mobile App platform Support: Android and IOS App Support for Cattle Artificial Insemination recording with other Cattle management information for better usability" to be included in the existing Technical specification.	No change
10.	Section-IV Modular OT	Specification as per bid document	In the technical specifications of MOT queries raised for 1.No. of Filters for AQS? 2. In Air handling Unit How much CFM used in OT? 3. OT light of how many Lux? Is it wall mounted or Portable. 4. Specifications of Anaesthesia machine. 5. Material quality of surgical equipment is TCTP, SS or German steel? 6. Autoclave is vertical or Horizontal? 7. How much epoxy flooring in mm? 8. Animal cage size? 9. Recovery cage size?	Revised Specification are attached at Annexure-I.
11.	Section-IV Gauze cloth	Specification as per bid document	Gauze cloth: - Thread count in TPI and txt of the yarn based on the given thread count.	No change

			- Is ISI certification is mandatory.	
12.	Section-IV Roller Bandage 10cm x3m & 15cm x3m gauze/cotton	Specification as per bid document	TPI count and exact dimension of 1 yam in Roller bandage 10cmX3 m, Roller bandage 10cmX3 m, Gauze than/Cotton	Item dimensions L x W x H 10 x 300 x 0.1 Centimeters may be ignored.
13.			Inclusion of distinct column for spare part, AMC/CMC in the tender document.	Separate column for CMC has been provided for 3 years beyond warranty period in the BoQ.

DIRECTOR


13.3
2

Annexure-I

Name of the Item	Revised Specification
Modular OT	In Air handling unit 5400 CFM used in OT. Minimum -10 HEPA filters for Air Quality standards OT light -ceiling mounted Surgical —SS304 Autoclave -vertical Epoxy flooring -3 mm Both Cage (Animal & Recovery) 8 x5x5 ft Specifications of Anesthesia Machine: Small animal anaesthesia machine-1500cc heavy duty "Coaxial" absorber, recessed. • Single large thumbscrew canister release. • 0.1 to 4lpm precision Oxygen flowmeter. • Rotary float and large scale. • Recessed flush valve to prevent accidental activation. • Quick connect fresh gas outlet for easy and secure switching between breathing circuits. • Compatibility for easy and quick installation of any model of Vaporizer. • Large Quick Lock Inspiration and Expiration domes for easy access and visibility. • Momentary closed feature incorporated into our pop-off valve. Circuit pressure gauge, easy to read and large. • Patient circuit pressure alarm audible. • Rotating bag mount. • Evacuation connector. • Frame finished with textured powder coat. • High stability 16 gauge construction.

8

- Pre tapped holes for easy installation of bain circuit adaptor and scavenger interface.
- Oxygen supply hose 9ft.
- Patient circuit with 0.5, 1, 2 and 3 and 5 Litre bags.
- High quality Selectatec Isoflurane Vaporizer.
- Scavenger interface with scavenging pipe.
- Bain circuit adaptor.
- Bain circuit (152 cm length).
- E tank oxygen pressure regulator with high pressure tubing.
- Two 10 litres O2 cylinders and one 20 litres O2 cylinder.
- Oxygen cylinder trolley-one.
- 5 different sizes of anesthesia masks for canines and felines.
- High quality Red rubber and plastic coaxial tubing.
- Endotracheal tubes disposable all ranges between 3 to 10mm – 2 each.
- Filler tube for isoflurane.
- Sodalime. Q.S.

 13.3
DIRECTOR